

Modello Laser	E-400
Serial Number RF	4PA11370194
Serial Number Sorgente	ET53633
Sample Rate:	2000
Sample Period:	00.01.9
Total Duration:	1.02.39

Data	08/03/2013 14.41
Sonda	0164L12R
Frequenza	30 KHz
Duty	60 %
Valore Minimo	520 W
Valore Medio	535 W
Valore Massimo	540 W
Deviazione Standard	3.2 W
Stabilità	1.19% +/-

