

Modello Laser	E-400
Serial Number RF	EA3180
Serial Number Sorgente	ET54249
Sample Rate:	2000
Sample Period:	00:00,3
Total Duration:	0:11:20

Data	15/05/2013 10:04
Sonda	0164L12R
Frequenza	50 KHZ
Duty	60 %
Valore Minimo	463 W
Valore Medio	494 W
Valore Massimo	510 W
Deviazione Standard	13,0 W
Stabilità	5,26% +/-

