

Modello Laser	G-150L
Serial Number RF	ND
Serial Number Sorgente	GB45749
Sample Rate:	2000
Sample Period:	00:00.3
Total Duration:	0:11:19

Data	28/01/2014 15:25
Sonda	0163L12R
Frequenza	1 KHz
Duty	60 %
Valore Minimo	201 W
Valore Medio	203 W
Valore Massimo	207 W
Deviazione Standard	1.9 W
Stabilità	1.88% +/-

