

Modello Laser	G-150L
Serial Number RF	ND
Serial Number Sorgente	GB45749
Sample Rate:	2000
Sample Period:	00:01.9
Total Duration:	1:02:18

Data	28/01/2014 14:20
Sonda	0163L12R
Frequenza	50 KHz
Duty	60 %
Valore Minimo	198 W
Valore Medio	202 W
Valore Massimo	207 W
Deviazione Standard	2.6 W
Stabilità	2.54% +/-

