

Modello Laser	<b>k-500</b>
Serial Number RF	<b>BMP011957-0002</b>
Serial Number Sorgente	<b>HB47087</b>
Sample Rate:	<b>2000</b>
Sample Period:	<b>00.01.9</b>
Total Duration:	<b>1.02.39</b>

Data	<b>21/03/2013 11.48</b>
Sonda	<b>0164L12R</b>
Frequenza	<b>30 KHz</b>
Duty	<b>60 %</b>
Valore Minimo	<b>557 W</b>
Valore Medio	<b>578 W</b>
Valore Massimo	<b>611 W</b>
Deviazione Standard	<b>8.6 W</b>
Stabilità	<b>2.98% +/-</b>

